



California Medical Device Recall Information



Recall Name

B. Braun Medical Recalls Dialog+ Hemodialysis Systems Due To a Potential Defect

Recall Date	Product Description	Recalling Firm	Recall Reason
4/1/16	Dialog+ Hemodialysis Systems	B. Braun Medical, Inc. Bethlehem, PA	<i>Potential for cracks in conductivity sensors that may allow air to enter into the solution.</i>
Recall Class	Product Identification	Distribution	Affected Dates
I	Suspect Model Numbers: <ul style="list-style-type: none">• 710200L• 710200K• 710200S• 710200U• 710500L• 710500K	CA , nationwide	Manufactured between: April 1, 2013 and July 3, 2013

FOR ADDITIONAL INFORMATION, PLEASE VISIT:

<http://www.fda.gov/MedicalDevices/Safety/ListofRecalls/ucm499127.htm>